

Survey: ZL1/22
month: March 2022
Participant-No: 9900540
date of issue: 12.03.22



Referenzinstitut für Bioanalytik

Všeobecná fakultní nemocnice v Praze
Centrální Laborator na ULBLD
Prof. MUDr. Zima, RNDr. Benáková
U Nemocnice 499/2, budova A 7, 2. patro
12808 PRAHA 2
Czech Republic



Survey directors

Prof. Dr. C. Knabbe
Prof. Dr. Dr. K.P. Kohse
Prof. Dr. M. Neumaier

Head of RfB

Dr. W. J. Geilenkeuser
Dr. A. Kessler

Bonn, 4. April 2022

Certificate

We confirm that you have participated in the survey for cell count in CSF.

You have met the requirements of the survey for the following analytes:

leucocytes (1) | erythrocytes (1)

This certificate is according to the current version of the RiliBÄK valid until the end of March 2023.

The number in parentheses characterizes the analytical method used.

The assignment of the number to the respective method and/or the respective instrument is to be taken from the total evaluation.

http://www.rfb-bio.de/certificates/?id=6605410&v_id=21
&version=1&doc=11031459034507A6E837C95156187

Survey: ZL1/22
 month: March 2022
 Participant-No: 9900540



Listing and Evaluation of all your results

Explanations

Certificate

A certificate is issued (given) for an analyte only if the basis for an evaluation of the accuracy is given by the guidelines of the German Medical Association and/or if an evaluation is possible in analogy to these guidelines (see comments on the evaluation) and, if both results for an analyte are within the given acceptance limits. (marked as '+' below C)

Certificate of participation

In the participation certificate all analytes which are included in the list on this page are listed.
 If all analytes are listed on the certificate no participation certificate is printed.

Legend:	C=Certification, M=No of method, R=your result, D=difference (R-T)	Certification:	+ = fulfilled (quotient D/Dmax <= 1.0)
	Dmax= maximum allowable amount of difference in measurement, partly given by annex 1 of the guideline of the BÄK (Dt. Ärzteblatt 111, Heft 38, 19.9.2014).		- = not fulfilled (quotient D/Dmax > 1.0)
	T = target value, either reference method value or assigned value. LL UL = lower resp. upper limit		± = certification cancelled because of technical and/or analytical reasons

	C	M	R	D/Dmax	T	LL	UL										
leucocytes [3]	+	1	A	192	-0.22	234	46.8	423									
			B	19.0	-0.19	22.5	4.50	40.5									
erythrocytes [3]	+	1	A	210	-0.25	263	52.5	474									
			B	26.0	-0.09	28.0	5.58	50.4									