

Survey: **CS2/25**
month: **August 2025**
Participant-No: **9900540**
valid from: **30.08.25**



page 1 of 3

Vseobecná fakultní nemocnice v Praze
Centrální Laborator na ULBLD
Prof. MUDr. Zima, RNDr. Benáková
U Nemocnice 499/2, budova A 7, 2. patro
12808 PRAHA 2
CZECH REPUBLIC



Bonn, 9. September 2025

Certificate

We confirm that you have participated in the survey for Immunosuppressives in blood.
This certificate is according to the current version of the RiliBÄK valid until the end of August 2026.
You have met the requirements of the survey for the following analytes:

cyclosporine (4)

A handwritten signature in blue ink, appearing to read 'Kohse, K. P.'.

Prof. Dr. Dr. K. P. Kohse
EQA scheme director

A handwritten signature in blue ink, appearing to read 'A. Kessler'.

Dr. Anja Kessler
Head of Reference Institute

A handwritten signature in blue ink, appearing to read 'M. Enders'.

Dr. Marika Enders
EQAS-Board

The number in parentheses characterizes the analytical method used.
The assignment of the number to the respective method and/or the respective instrument
is to be taken from the total evaluation.

This certificate is valid in conjunction with the final report dated 09.09.25.
This is available for download in the RfB-Online system.

Survey: **CS2/25**
 month: **August 2025**
 Participant-No: **9900540**



page 2 of 3

Listing and Evaluation of all your results

Explanations

Certificate



A certificate is issued (given) for an analyte only if the basis for an evaluation of the accuracy is given by the guidelines of the German Medical Association and/or if an evaluation is possible in analogy to these guidelines (see comments on the evaluation) and, if both results for an analyte are within the given acceptance limits. (marked as '+' below C)

Certificate of participation



In the participation certificate all analytes which are included in the list on this page are listed.
 If all analytes are listed on the certificate no participation certificate is printed.

Legend:	C=Certification, M=No of method, R=your result, D=difference (R-T)	Certification:	+ = fulfilled (quotient D/Dmax <= 1.0)
	Dmax= maximum allowable amount of difference in measurement, partly given by annex 1 of the guideline of the BÄK.		- = not fulfilled (quotient D/Dmax > 1.0
	T = target value, either reference method value or assigned value, LL UL = lower resp. upper limit		± = certification cancelled because of technical and/or analytical reasons

	C	M	R	D/Dmax	T	LL	UL								
cyclosporine [µg/l]	+	4	A 40.0	-0.53	47.6	33.3	61.9								
			B 374	0.34	339	237	441								
cyclosporine [µg/l]	+	4	C 30.0	-0.36	33.7	23.5	43.9								
			D 83.0	-0.28	90.7	63.4	118								



Other kits (number):
2-38(1), 2-99(2), 4-40(1), 4-44(1), 4-99(1), 4-335(2), 5-135(1), 6-27(1), 6-39(1), 6-134(2), 7-30(1), 7-43(1), 7-75(1), 8-47(1), 8-335(1),



Other kits (number):
2-38(1), 2-99(2), 4-40(1), 4-44(1), 4-99(1), 4-335(2), 5-135(1), 6-27(1), 6-39(1), 6-134(2), 7-30(1), 7-43(1), 7-75(1), 8-47(1), 8-335(1),