

Survey: ZY3/26
month: August 2026
Participant-No: 9900540
valid from: 08.08.26



page 1 of 3

Všeobecná fakultní nemocnice v Praze
Centrální laborator na ULBLD
Prof. MUDr. Zima, RNDr. Benáková
U Nemocnice 499/2, budova A 7, 2. patro
12808 PRAHA 2
CZECH REPUBLIC



Bonn, 24. August 2026

Certificate

We confirm that you have participated in the survey for cytokines.
This certificate is according to the current version of the RiliBÄK valid until the end of February 2027.

You have met the requirements of the survey for the following analytes:

IL-6 (3) | PCT (3)

Handwritten signature of Prof. Dr. Dr. K. P. Kohse in blue ink.

Prof. Dr. Dr. K. P. Kohse
EQA scheme director

Handwritten signature of Dr. Anja Kessler in blue ink.

Dr. Anja Kessler
Head of Reference Institute

Handwritten signature of Dr. Marika Enders in blue ink.

Dr. Marika Enders
EQAS-Board

The number in parentheses characterizes the analytical method used.
The assignment of the number to the respective method and/or the respective instrument
is to be taken from the total evaluation.

This certificate is valid in conjunction with the final report dated 21.08.26.
This is available for download in the RfB-Online system.

Survey: **ZY3/26**
 month: **August 2026**
 Participant-No: **9900540**



Listing and Evaluation of all your results

Explanations

Certificate

A certificate is issued (given) for an analyte only if the basis for an evaluation of the accuracy is given by the guidelines of the German Medical Association and/or if an evaluation is possible in analogy to these guidelines (see comments on the evaluation) and, if both results for an analyte are within the given acceptance limits. (marked as '+' below C)

Participation certificate

In the participation certificate all analytes which are included in the list on this page are listed.
 If all analytes are listed on the certificate no participation certificate is printed.

Legend:	C=Certification, M=No of method, R=your result, D=difference (R-T) Dmax= maximum allowable amount of difference in measurement, partly given by annex 1 of the guideline of the BÄK. T = target value, either reference method value or assigned value, LL UL = lower resp. upper limit	Certification:	+ = fulfilled (quotient D/Dmax <= 1.0) - = not fulfilled (quotient D/Dmax > 1.0 ± = certification cancelled because of technical and/or analytical reasons
---------	---	----------------	--

	C	M	R	D/Dmax	T	LL	UL								
IL-6	+	3	A	239	0.10	232	162	302							
[ng/l]			B	84.5	-0.30	92.9	65.0	121							
PCT	+	3	A	0.450	0.15	0.430	0.301	0.559							
[µg/l]			B	5.42	0.13	5.21	3.64	6.78							

